
Midmark M11 - v1.0.5
Total Cycles: 3174

Name: RITTER2
Sterilizer ID: V2313228
12/01/2023 09:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.9 F 0.0
2:00 69.2 F 0.0
4:00 72.1 F 0.2
6:00 97.5 F 1.0
8:00 138.0 F 3.4
10:00 161.2 F 6.6
12:00 178.0 F 10.0
14:00 196.0 F 13.5
16:00 209.6 F 5.6
18:00 221.2 F 9.3
20:00 232.1 F 13.4
22:00 234.2 F 9.7
24:00 238.0 F 10.2
26:00 244.4 F 13.1
28:00 255.3 F 18.8
30:00 265.2 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.5 F 29.9
2:00 271.5 F 29.1
3:00 272.2 F 30.0
4:00 272.1 F 29.5

Min 271.2 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.8 F 0.0
5:00 215.2 F 0.0
10:00 200.9 F 0.0
15:00 189.8 F 0.0
20:00 188.4 F 0.0
25:00 187.7 F 0.0
30:00 191.7 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:59
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:10:13
CYCLE COMPLETE

12/01/2023 11:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____