
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004124
12:05 27/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:11
PRESSURIZING 2:15
STERILIZING 3:11
136.0 C 220kPa 3:11
136.4 C 224kPa 3:41
136.1 C 223kPa 4:11
136.2 C 223kPa 4:41
136.2 C 221kPa 5:11
136.4 C 221kPa 5:41
136.4 C 225kPa 6:11
136.4 C 219kPa 6:41
136.4 C 219kPa 6:41
VENTING 6:41
AIR DRYING 8:43
CYCLE COMPLETE 9:00

Digital Signature #
760C1DFEBAC6FD3A
