

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000030  
11:34 09/01/2020  
WRAPPED  
135 C for 10 min.  
CYCLE START 0:00  
CONDITIONING 3:12  
PRESSURIZING 5:20  
STERILIZING 6:14  
136.0 C 220kPa 6:14  
136.1 C 222kPa 6:44  
136.4 C 222kPa 7:14  
136.1 C 222kPa 7:44  
136.3 C 224kPa 8:14  
136.2 C 223kPa 8:44  
136.3 C 224kPa 9:14  
136.4 C 224kPa 9:44  
136.5 C 221kPa 10:14  
136.4 C 224kPa 10:44  
136.4 C 224kPa 11:14  
136.3 C 223kPa 11:44  
136.1 C 224kPa 12:14  
136.2 C 223kPa 12:44  
136.2 C 222kPa 13:14  
136.5 C 224kPa 13:44  
136.4 C 220kPa 14:14  
136.3 C 223kPa 14:44  
136.3 C 222kPa 15:14  
136.2 C 219kPa 15:44  
136.5 C 225kPa 16:14  
136.5 C 225kPa 16:14  
VENTING 16:14  
AIR DRYING 17:11  
CYCLE COMPLETE 17:15  
-----

Digital Signature #

857723AF2D8EF800

-----