
Midmark M11 - v1.0.5
Total Cycles: 3968

Name: RITTER2
Sterilizer ID: V2313228
09/23/2024 01:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 95.7 F 0.0
2:00 98.0 F 0.0
4:00 121.8 F 0.1
6:00 162.6 F 0.9
8:00 184.4 F 4.2
10:00 193.0 F 8.3
12:00 211.3 F 12.3
14:00 222.7 F 6.8
16:00 231.1 F 11.5
18:00 233.2 F 8.6
20:00 241.0 F 11.1
22:00 245.2 F 13.3
24:00 257.0 F 19.6
26:00 267.8 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.5 F 29.7
2:00 271.5 F 28.9
3:00 272.3 F 29.8
4:00 272.1 F 29.3

Min 271.2 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 205.9 F 0.0
10:00 196.8 F 0.0
15:00 185.1 F 0.0
20:00 181.4 F 0.0
25:00 180.0 F 0.0
30:00 179.4 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 27:21
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:05:48
CYCLE COMPLETE

09/23/2024 02:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____