

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1152

Name: RITTER2  
Sterilizer ID: V2313228  
06/01/2022 12:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 144.8 F 0.0  
2:00 145.0 F 2.2  
4:00 157.0 F 5.4  
6:00 182.4 F 9.7  
8:00 208.3 F 6.6  
10:00 221.7 F 7.9  
12:00 234.0 F 12.7  
14:00 235.2 F 9.9  
16:00 236.7 F 9.2  
18:00 247.4 F 14.6  
20:00 259.1 F 21.1  
22:00 269.7 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.1  
1:00 271.2 F 29.3  
2:00 272.5 F 30.1  
3:00 272.7 F 29.8  
4:00 271.9 F 29.3

Min 271.2 F 29.0  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.6 F 0.0  
5:00 206.0 F 0.0  
10:00 209.9 F 0.0  
15:00 205.5 F 0.0  
20:00 205.2 F 0.0  
25:00 207.2 F 0.0  
30:00 207.7 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 22:55  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:01:04  
CYCLE COMPLETE

06/01/2022 01:31 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_