
Midmark M11 - v1.0.0
Total Cycles: 4453

Name: RITTER 1
Sterilizer ID: V2178198
04/12/2024 01:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.4 F 0.0
2:00 113.3 F 1.1
4:00 132.8 F 3.4
6:00 163.2 F 6.1
8:00 186.3 F 9.9
10:00 208.5 F 9.8
12:00 211.4 F 5.3
14:00 221.7 F 9.0
16:00 232.2 F 13.0
18:00 233.1 F 9.1
20:00 240.6 F 11.5
22:00 242.6 F 12.2
24:00 254.0 F 17.9
26:00 264.1 F 24.0
28:00 271.9 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.6 F 29.7
2:00 272.0 F 29.1
3:00 271.3 F 28.9
4:00 272.5 F 29.7

Min 271.2 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.2 F 0.0
5:00 209.5 F 0.0
10:00 220.1 F 0.0
15:00 227.0 F 0.0
20:00 232.9 F 0.0
25:00 228.7 F 0.0
30:00 228.3 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 28:11
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:07:01
CYCLE COMPLETE

04/12/2024 02:33 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____