
Midmark M11 - v1.0.0
Total Cycles: 4366

Name: RITTER 1
Sterilizer ID: V2178198
03/07/2024 02:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.9 F 0.0
2:00 128.5 F 1.6
4:00 141.5 F 4.5
6:00 167.5 F 7.8
8:00 185.7 F 11.8
10:00 212.2 F 5.2
12:00 221.2 F 9.5
14:00 226.5 F 10.5
16:00 236.0 F 10.1
18:00 235.6 F 8.7
20:00 246.1 F 13.5
22:00 256.6 F 19.5
24:00 267.3 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 273.0 F 29.6
2:00 272.1 F 29.1
3:00 271.2 F 28.8
4:00 272.4 F 29.6

Min 271.1 F 28.7
Max 273.1 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 223.1 F 0.0
10:00 230.3 F 0.0
15:00 227.2 F 0.0
20:00 224.0 F 0.0
25:00 222.8 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 25:33
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:04:23
CYCLE COMPLETE

03/07/2024 03:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____