
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006440
9:32 11/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:08
STERILIZING 4:11
136.5 C 221kPa 4:11
136.4 C 225kPa 4:41
136.4 C 221kPa 5:11
136.5 C 224kPa 5:41
136.0 C 220kPa 6:11
136.4 C 223kPa 6:41
136.4 C 223kPa 7:11
136.4 C 223kPa 7:41
136.4 C 223kPa 7:41
VENTING 7:41
AIR DRYING 8:33
CYCLE COMPLETE 8:40

Digital Signature #
F401899B9E3FAC17
