
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007018
8:33 01/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:05
PRESSURIZING 2:08
STERILIZING 3:11
136.4 C 221kPa 3:11
136.4 C 222kPa 3:41
136.4 C 222kPa 4:11
136.2 C 222kPa 4:41
136.4 C 223kPa 5:11
136.2 C 220kPa 5:41
135.9 C 221kPa 6:11
135.9 C 220kPa 6:41
135.9 C 220kPa 6:41
VENTING 6:41
AIR DRYING 7:31
CYCLE COMPLETE 18:27

Digital Signature #
4C5DFE46E7FBFC9D
