
Midmark M11 - v1.0.5
Total Cycles: 2762

Name: RITTER2
Sterilizer ID: V2313228
08/14/2023 03:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.0 F 0.0
2:00 174.9 F 0.0
4:00 200.2 F 0.9
6:00 214.1 F 5.2
8:00 230.3 F 9.6
10:00 233.8 F 9.9
12:00 240.2 F 11.7
14:00 239.0 F 10.3
16:00 237.4 F 9.3
18:00 251.1 F 16.0
20:00 262.7 F 22.7
22:00 272.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 271.8 F 28.7
2:00 272.1 F 29.0
3:00 272.3 F 29.4
4:00 271.4 F 28.7

Min 271.4 F 28.6
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.6 F 0.0
5:00 238.0 F 0.0
10:00 221.3 F 0.0
15:00 185.8 F 0.0
20:00 179.8 F 0.0
25:00 176.6 F 0.0
30:00 175.0 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 22:16
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:01:31
CYCLE COMPLETE

08/14/2023 04:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____