

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3400

Name: RITTER2  
Sterilizer ID: V2313228  
02/21/2024 07:06 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.4 F 0.0  
2:00 171.1 F 2.6  
4:00 190.7 F 7.0  
6:00 208.7 F 12.1  
8:00 222.3 F 6.4  
10:00 234.1 F 11.6  
12:00 235.3 F 9.3  
14:00 238.1 F 9.5  
16:00 249.5 F 15.3  
18:00 261.9 F 22.5  
20:00 272.6 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.9  
1:00 272.7 F 29.3  
2:00 272.0 F 29.4  
3:00 272.4 F 29.2  
4:00 271.8 F 29.2

Min 271.5 F 28.7  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.1 F 0.0  
5:00 221.1 F 0.0  
10:00 213.4 F 0.0  
15:00 211.3 F 0.0  
20:00 208.8 F 0.0  
25:00 207.6 F 0.0  
30:00 208.7 F 0.0  
Drying Complete

FILLING: 2:02  
HEATING: 20:18  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 00:59:48  
CYCLE COMPLETE

02/21/2024 08:05 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_