
Midmark M11 - v1.0.0
Total Cycles: 724

Name: RITTER 1
Sterilizer ID: V2178198
03/17/2021 10:06 AM

BEGIN PACKS CYCLE
Temp: 250 Degrees F
Time: 30 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.7 F 0.0
2:00 181.9 F 2.4
4:00 183.4 F 6.5
6:00 207.6 F 2.3
8:00 222.1 F 6.9
10:00 225.6 F 5.2
12:00 226.4 F 4.5
14:00 239.7 F 10.2
16:00 251.6 F 16.0

STERILIZING
mm:ss Degrees PSI
0:00 251.5 F 15.9
1:00 252.0 F 16.1
2:00 252.7 F 16.4
3:00 251.5 F 16.0
4:00 252.3 F 16.3
5:00 251.7 F 16.1
6:00 252.5 F 16.3
7:00 251.5 F 15.9
8:00 252.8 F 16.5
9:00 252.1 F 16.2
10:00 251.5 F 15.9
11:00 252.8 F 16.5
12:00 252.3 F 16.3
13:00 251.8 F 16.0
14:00 251.7 F 16.1
15:00 253.0 F 16.5
16:00 252.4 F 16.3
17:00 251.8 F 16.1
18:00 251.5 F 15.9
19:00 252.4 F 16.5
20:00 252.7 F 16.5
21:00 252.3 F 16.3
22:00 251.8 F 16.1
23:00 251.4 F 15.9
24:00 252.0 F 16.3
25:00 252.9 F 16.5
26:00 252.5 F 16.4
27:00 252.2 F 16.2
28:00 251.8 F 16.0
29:00 251.3 F 15.9
30:00 252.0 F 16.4

Min 251.3 F 15.8
Max 253.0 F 16.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.3 F 0.0
5:00 206.4 F 0.0
10:00 198.1 F 0.0
15:00 197.3 F 0.0
20:00 194.7 F 0.0
25:00 192.8 F 0.0
30:00 192.6 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 16:31

STERILIZING: 30:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:20:52
CYCLE COMPLETE

03/17/2021 11:51 AM
ACCEPTED: X
REJECTED:
Operator: U01