
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001506
10:51 10/15/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:54
PRESSURIZING 6:40
STERILIZING 8:16
136.0 C 220kPa 8:16
135.9 C 217kPa 8:46
136.3 C 222kPa 9:16
135.7 C 219kPa 9:46
136.3 C 221kPa 10:16
136.0 C 218kPa 10:46
136.4 C 223kPa 11:16
136.4 C 220kPa 11:46
VENTING 11:46
AIR DRYING 12:46
CYCLE COMPLETE 42:46

Digital Signature #
C2FC2A8163E87BE6
