
Midmark M11 - v1.0.0
Total Cycles: 3028

Name: RITTER 1
Sterilizer ID: V2178198
03/22/2023 02:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.0 F 0.0
2:00 188.1 F 3.9
4:00 201.1 F 8.6
6:00 223.4 F 9.9
8:00 231.6 F 9.2
10:00 238.9 F 9.8
12:00 241.8 F 12.1
14:00 240.4 F 10.6
16:00 253.6 F 17.2
18:00 265.1 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.5 F 28.9
2:00 271.4 F 28.5
3:00 272.5 F 29.0
4:00 271.4 F 28.3

Min 271.3 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.6 F 0.0
5:00 210.1 F 0.0
10:00 218.9 F 0.0
15:00 226.0 F 0.0
20:00 235.8 F 0.0
25:00 228.3 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 19:50
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:41
CYCLE COMPLETE

03/22/2023 03:30 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____