
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 005054
11:59 02/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:33
STERILIZING 3:26
136.0 C 220kPa 3:26
136.3 C 223kPa 3:56
136.1 C 218kPa 4:26
136.1 C 219kPa 4:56
136.1 C 221kPa 5:26
136.3 C 224kPa 5:56
136.4 C 224kPa 6:26
136.4 C 220kPa 6:56
136.4 C 220kPa 6:56
VENTING 6:56
AIR DRYING 7:54
CYCLE COMPLETE 11:14

Digital Signature #
848C57C6519F48D1
