
Midmark M11 - v1.0.5
Total Cycles: 1751

Name: RITTER2
Sterilizer ID: V2313228
11/21/2022 12:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.9 F 0.0
2:00 187.0 F 1.4
4:00 194.9 F 6.3
6:00 215.1 F 11.8
8:00 227.0 F 7.9
10:00 238.2 F 11.9
12:00 243.0 F 12.6
14:00 243.6 F 12.3
16:00 257.2 F 19.7
18:00 269.4 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.0
1:00 272.2 F 29.7
2:00 272.2 F 29.3
3:00 271.5 F 29.2
4:00 272.7 F 29.8

Min 271.3 F 28.8
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.7 F 0.0
5:00 225.3 F 0.0
10:00 233.7 F 0.0
15:00 230.4 F 0.0
20:00 222.4 F 0.0
25:00 218.7 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 18:52
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:27
CYCLE COMPLETE

11/21/2022 01:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____