
Midmark M11 - v1.0.5
Total Cycles: 4425

Name: RITTER2
Sterilizer ID: V2313228
03/13/2025 12:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.2 F 0.0
2:00 137.3 F 1.4
4:00 148.6 F 4.2
6:00 179.8 F 8.1
8:00 199.8 F 11.8
10:00 217.5 F 5.4
12:00 226.1 F 9.5
14:00 237.0 F 11.5
16:00 237.7 F 11.0
18:00 235.7 F 8.9
20:00 248.6 F 15.2
22:00 259.6 F 21.3
24:00 269.3 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 30.0
1:00 271.8 F 30.2
2:00 271.9 F 30.4
3:00 271.8 F 30.4
4:00 271.8 F 30.4

Min 271.7 F 29.9
Max 272.7 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 205.4 F 0.0
10:00 211.7 F 0.0
15:00 205.7 F 0.0
20:00 200.5 F 0.0
25:00 197.5 F 0.0
30:00 194.5 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 25:38
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:04:35
CYCLE COMPLETE

03/13/2025 01:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____