

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2844

Name: RITTER 1  
Sterilizer ID: V2178198  
02/01/2023 04:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.7 F 0.0  
2:00 178.7 F 3.8  
4:00 203.2 F 8.1  
6:00 207.6 F 13.1  
8:00 222.8 F 6.7  
10:00 233.9 F 11.7  
12:00 234.3 F 8.3  
14:00 241.0 F 11.0  
16:00 244.5 F 12.7  
18:00 256.9 F 19.3  
20:00 268.1 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 272.5 F 29.5  
2:00 272.3 F 29.1  
3:00 271.6 F 28.6  
4:00 271.7 F 29.0  
  
Min 271.3 F 28.5  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.3 F 0.0  
5:00 213.4 F 0.0  
10:00 230.8 F 0.0  
15:00 234.6 F 0.0  
20:00 225.1 F 0.0  
25:00 225.9 F 0.0  
30:00 227.9 F 0.0  
Drying Complete

FILLING: 1:16  
HEATING: 21:22  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:59:55  
CYCLE COMPLETE

02/01/2023 05:02 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_