
Midmark M11 - v1.0.5
Total Cycles: 3835

Name: RITTER2
Sterilizer ID: V2313228
08/07/2024 03:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 100.5 F 0.0
2:00 101.8 F 0.0
4:00 111.3 F 0.6
6:00 154.0 F 3.6
8:00 173.9 F 7.0
10:00 212.7 F 11.0
12:00 217.3 F 5.4
14:00 224.2 F 9.0
16:00 235.9 F 13.6
18:00 235.2 F 10.6
20:00 235.4 F 8.4
22:00 247.7 F 14.7
24:00 258.6 F 20.8
26:00 268.7 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 271.4 F 29.3
2:00 271.5 F 29.3
3:00 272.1 F 30.5
4:00 272.6 F 30.1

Min 271.0 F 29.3
Max 273.7 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.8 F 0.0
5:00 226.6 F 0.0
10:00 222.9 F 0.0
15:00 220.4 F 0.0
20:00 201.3 F 0.0
25:00 202.3 F 0.0
30:00 198.2 F 0.0

Drying Complete

FILLING: 0:59
HEATING: 27:05
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:05:42
CYCLE COMPLETE

08/07/2024 04:22 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____