
Midmark M11 - v1.0.5
Total Cycles: 1465

Name: RITTER2
Sterilizer ID: V2313228
09/07/2022 03:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.9 F 0.0
2:00 165.1 F 3.3
4:00 182.3 F 7.8
6:00 204.7 F 13.3
8:00 221.9 F 8.0
10:00 235.9 F 13.5
12:00 238.9 F 11.2
14:00 238.8 F 10.2
16:00 252.9 F 17.1
18:00 265.0 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.6 F 29.4
2:00 271.5 F 28.6
3:00 272.7 F 29.6
4:00 272.2 F 29.0

Min 271.4 F 28.6
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.4 F 0.0
5:00 205.1 F 0.0
10:00 220.8 F 0.0
15:00 226.7 F 0.0
20:00 227.3 F 0.0
25:00 219.3 F 0.0
30:00 216.9 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 19:47
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:25
CYCLE COMPLETE

09/07/2022 04:20 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____