

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000838  
14:25 05/01/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:46  
PRESSURIZING 2:56  
STERILIZING 4:16  
136.0 C 220kPa 4:16  
136.4 C 233kPa 4:47  
136.4 C 233kPa 5:17  
136.1 C 229kPa 5:47  
136.4 C 233kPa 6:17  
136.3 C 233kPa 6:47  
136.4 C 233kPa 7:17  
136.2 C 231kPa 7:47  
136.2 C 231kPa 7:47  
VENTING 7:47  
AIR DRYING 8:37  
CYCLE COMPLETE 10:50  
-----

Digital Signature #  
D538032E0E279952  
-----