
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003750
11:47 01/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:31
STERILIZING 3:31
136.3 C 220kPa 3:31
136.4 C 223kPa 4:01
136.4 C 224kPa 4:31
136.4 C 224kPa 5:01
136.3 C 225kPa 5:31
136.4 C 224kPa 6:01
136.0 C 222kPa 6:31
136.4 C 219kPa 7:01
136.4 C 219kPa 7:01
VENTING 7:01
AIR DRYING 8:52
CYCLE COMPLETE 9:13

Digital Signature #
FA0CDC294EB1140B
