
Midmark M11 - v1.0.5
Total Cycles: 3515

Name: RITTER2
Sterilizer ID: V2313228
04/11/2024 08:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.6 F 0.0
2:00 69.8 F 0.0
4:00 82.1 F 0.0
6:00 116.9 F 0.1
8:00 146.9 F 0.5
10:00 172.1 F 3.5
12:00 197.1 F 7.0
14:00 209.5 F 10.5
16:00 224.4 F 7.7
18:00 224.7 F 7.9
20:00 237.6 F 12.0
22:00 234.1 F 8.6
24:00 241.9 F 11.8
26:00 244.1 F 12.6
28:00 255.6 F 18.7
30:00 265.9 F 25.0
32:00 272.5 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 272.4 F 29.1
2:00 272.8 F 29.4
3:00 272.6 F 29.7
4:00 271.7 F 29.2

Min 271.7 F 29.0
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.8 F 0.0
5:00 225.0 F 0.0
10:00 179.0 F 0.0
15:00 174.7 F 0.0
20:00 171.5 F 0.0
25:00 170.6 F 0.0
30:00 169.9 F 0.0

Drying Complete

FILLING: 0:48
HEATING: 32:25
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:10:31
CYCLE COMPLETE

04/11/2024 09:40 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____