

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3368

Name: RITTER2  
Sterilizer ID: V2313228  
02/08/2024 09:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.7 F 0.0  
2:00 71.0 F 0.0  
4:00 76.5 F 0.0  
6:00 105.3 F 1.2  
8:00 145.8 F 4.1  
10:00 171.7 F 7.2  
12:00 195.5 F 10.7  
14:00 207.3 F 5.6  
16:00 217.5 F 7.7  
18:00 229.3 F 11.7  
20:00 228.7 F 7.4  
22:00 240.9 F 12.6  
24:00 239.2 F 10.5  
26:00 251.4 F 16.5  
28:00 262.1 F 22.8  
30:00 272.4 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.9  
1:00 272.7 F 29.7  
2:00 271.5 F 28.8  
3:00 272.3 F 29.8  
4:00 272.3 F 29.2

Min 271.0 F 28.7  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.7 F 0.0  
5:00 224.6 F 0.0  
10:00 191.0 F 0.0  
15:00 188.2 F 0.0  
20:00 185.6 F 0.0  
25:00 183.3 F 0.0  
30:00 182.8 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:29  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:08:29  
CYCLE COMPLETE

02/08/2024 11:01 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_