
Midmark M11 - v1.0.0
Total Cycles: 4370

Name: RITTER 1
Sterilizer ID: V2178198
03/08/2024 11:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.8 F 0.0
2:00 161.0 F 1.9
4:00 165.5 F 5.0
6:00 195.8 F 9.0
8:00 217.2 F 11.5
10:00 223.4 F 6.7
12:00 234.7 F 10.9
14:00 230.2 F 6.8
16:00 243.8 F 12.2
18:00 239.9 F 10.1
20:00 251.8 F 16.1
22:00 262.7 F 22.2
24:00 272.5 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.6 F 28.7
2:00 272.2 F 29.1
3:00 271.5 F 28.4
4:00 272.4 F 29.0

Min 271.1 F 28.3
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.4 F 0.0
5:00 238.4 F 0.0
10:00 230.4 F 0.0
15:00 211.8 F 0.0
20:00 192.5 F 0.0
25:00 209.0 F 0.0
30:00 214.4 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 25:11
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:03:54
CYCLE COMPLETE

03/08/2024 12:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____