
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000925
17:51 10/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:10
PRESSURIZING 4:10
STERILIZING 5:52
136.4 C 231kPa 5:52
136.4 C 233kPa 6:22
136.3 C 232kPa 6:52
136.1 C 231kPa 7:22
136.1 C 231kPa 7:52
136.4 C 233kPa 8:22
136.3 C 232kPa 8:52
136.4 C 233kPa 9:22
136.4 C 233kPa 9:22
VENTING 9:22
AIR DRYING 10:27
CYCLE COMPLETE 17:33

Digital Signature #
804B9809E4636FD9
