
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001496
15:44 04/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:38
STERILIZING 3:46
136.5 C 224kPa 3:46
136.4 C 225kPa 4:16
136.4 C 224kPa 4:46
136.4 C 222kPa 5:16
136.4 C 220kPa 5:46
136.4 C 223kPa 6:16
136.5 C 223kPa 6:46
136.2 C 223kPa 7:16
136.2 C 223kPa 7:16
VENTING 7:16
AIR DRYING 8:05
CYCLE COMPLETE 14:24

Digital Signature #
64C2B01F165B4BDE
