

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3657

Name: RITTER2  
Sterilizer ID: V2313228  
05/29/2024 03:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 82.4 F 0.0  
2:00 82.8 F 0.0  
4:00 88.0 F 0.3  
6:00 108.1 F 1.7  
8:00 143.3 F 4.3  
10:00 185.5 F 7.2  
12:00 196.0 F 10.6  
14:00 209.5 F 6.7  
16:00 215.4 F 6.8  
18:00 226.5 F 10.4  
20:00 233.9 F 9.0  
22:00 237.0 F 11.0  
24:00 236.6 F 8.3  
26:00 246.5 F 14.2  
28:00 256.4 F 19.7  
30:00 266.0 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.3  
1:00 271.2 F 29.5  
2:00 272.4 F 30.1  
3:00 272.4 F 29.8  
4:00 271.5 F 29.4  
  
Min 271.0 F 29.1  
Max 272.7 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.7 F 0.0  
5:00 208.4 F 0.0  
10:00 200.5 F 0.0  
15:00 199.0 F 0.0  
20:00 194.8 F 0.0  
25:00 210.3 F 0.0  
30:00 216.7 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 31:48  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:10:13  
CYCLE COMPLETE

05/29/2024 04:12 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_