
Midmark M11 - v1.0.0
Total Cycles: 5017

Name: RITTER 1
Sterilizer ID: V2178198
10/23/2024 12:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.8 F 0.0
2:00 124.0 F 0.0
4:00 144.6 F 0.3
6:00 155.0 F 0.9
8:00 177.7 F 3.5
10:00 193.7 F 6.5
12:00 211.7 F 10.2
14:00 227.8 F 8.4
16:00 225.8 F 7.8
18:00 235.4 F 12.1
20:00 234.8 F 8.8
22:00 243.1 F 11.5
24:00 244.0 F 12.4
26:00 255.1 F 18.2
28:00 265.2 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 271.5 F 28.7
2:00 272.6 F 29.2
3:00 272.0 F 28.8
4:00 271.2 F 28.4

Min 271.2 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 209.9 F 0.0
10:00 196.3 F 0.0
15:00 208.8 F 0.0
20:00 224.5 F 0.0
25:00 242.0 F 0.0
30:00 243.9 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 29:52
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:08:24
CYCLE COMPLETE

10/23/2024 01:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____