
Midmark M11 - v1.0.5
Total Cycles: 2693

Name: RITTER2
Sterilizer ID: V2313228
07/24/2023 07:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.5 F 0.0
2:00 182.5 F 3.8
4:00 201.7 F 9.3
6:00 219.3 F 7.7
8:00 231.3 F 10.6
10:00 232.2 F 8.2
12:00 239.5 F 10.3
14:00 247.8 F 14.5
16:00 260.6 F 21.7
18:00 272.1 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.3 F 29.2
2:00 272.5 F 29.8
3:00 271.9 F 28.9
4:00 272.7 F 29.6

Min 271.6 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.4 F 0.0
5:00 227.3 F 0.0
10:00 242.0 F 0.0
15:00 233.8 F 0.0
20:00 223.7 F 0.0
25:00 220.6 F 0.0
30:00 218.6 F 0.0
Drying Complete

FILLING: 2:41
HEATING: 18:28
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:17
CYCLE COMPLETE

07/24/2023 08:32 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____