
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001987
7:47 09/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:29
PRESSURIZING 3:30
STERILIZING 4:36
136.0 C 220kPa 4:36
136.2 C 223kPa 5:06
136.1 C 220kPa 5:36
136.3 C 223kPa 6:06
136.1 C 219kPa 6:36
136.4 C 223kPa 7:06
136.3 C 222kPa 7:36
136.4 C 224kPa 8:06
136.4 C 224kPa 8:06
VENTING 8:06
AIR DRYING 9:02
CYCLE COMPLETE 39:02

Digital Signature #
F804ABF793137C6C
