

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2608

Name: RITTER2  
Sterilizer ID: V2313228  
07/05/2023 09:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.6 F 0.0  
2:00 192.1 F 3.2  
4:00 196.7 F 8.1  
6:00 214.0 F 13.2  
8:00 228.3 F 9.1  
10:00 235.1 F 9.8  
12:00 242.7 F 12.9  
14:00 242.3 F 12.0  
16:00 255.6 F 18.9  
18:00 267.0 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.5 F 29.6  
2:00 271.4 F 29.0  
3:00 272.4 F 29.7  
4:00 271.9 F 29.1  
  
Min 271.4 F 28.9  
Max 272.5 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.5 F 0.0  
5:00 209.5 F 0.0  
10:00 218.5 F 0.0  
15:00 218.2 F 0.0  
20:00 216.5 F 0.0  
25:00 214.8 F 0.0  
30:00 212.9 F 0.0  
Drying Complete

FILLING: 2:04  
HEATING: 19:29  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:42  
CYCLE COMPLETE

07/05/2023 10:56 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_