

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1467

Name: RITTER 1  
Sterilizer ID: V2178198  
12/16/2021 11:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.7 F 0.0  
2:00 189.8 F 3.0  
4:00 208.3 F 7.8  
6:00 219.1 F 10.0  
8:00 230.1 F 9.0  
10:00 240.3 F 10.1  
12:00 243.4 F 12.8  
14:00 243.0 F 11.9  
16:00 256.7 F 18.9  
18:00 268.5 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.3 F 28.8  
2:00 271.4 F 28.5  
3:00 272.3 F 28.9  
4:00 271.2 F 28.2  
  
Min 271.2 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.4 F 0.0  
5:00 218.0 F 0.0  
10:00 220.4 F 0.0  
15:00 220.2 F 0.0  
Dry Cycle Aborted

FILLING: 2:03  
HEATING: 19:09  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 19:03  
TOTAL CYCLE: 00:47:24

12/16/2021 12:54 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02