
Midmark M11 - v1.0.0
Total Cycles: 2866

Name: RITTER 1
Sterilizer ID: V2178198
02/07/2023 08:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.0 F 0.0
2:00 80.3 F 0.0
4:00 109.0 F 0.5
6:00 123.6 F 2.7
8:00 148.3 F 5.5
10:00 171.8 F 8.8
12:00 195.2 F 12.8
14:00 209.3 F 4.6
16:00 218.2 F 8.4
18:00 229.6 F 12.5
20:00 231.8 F 8.1
22:00 243.6 F 13.3
24:00 241.8 F 11.3
26:00 254.1 F 17.5
28:00 264.8 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.4 F 29.1
2:00 271.4 F 28.5
3:00 272.4 F 29.1
4:00 271.4 F 28.3

Min 271.3 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.6 F 0.0
5:00 224.4 F 0.0
10:00 225.4 F 0.0
15:00 222.9 F 0.0
20:00 222.5 F 0.0
25:00 230.9 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 29:54
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:07:52
CYCLE COMPLETE

02/07/2023 09:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____