

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1133

Name: RITTER 1  
Sterilizer ID: V2178198  
08/31/2021 07:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.5 F 0.0  
2:00 77.2 F 0.0  
4:00 110.3 F 0.0  
6:00 126.4 F 0.0  
8:00 162.3 F 0.0  
10:00 183.4 F 0.1  
12:00 202.4 F 0.2  
14:00 213.1 F 2.8  
16:00 220.7 F 5.4  
18:00 227.3 F 8.2  
20:00 236.2 F 11.8  
22:00 232.7 F 7.9  
24:00 244.4 F 13.1  
26:00 242.2 F 11.4  
28:00 239.8 F 10.1  
30:00 252.6 F 16.4  
32:00 263.6 F 22.8  
34:00 272.3 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.0  
1:00 272.6 F 28.8  
2:00 272.1 F 28.3  
3:00 271.5 F 28.1  
4:00 272.6 F 28.9  
  
Min 271.4 F 27.9  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.0 F 0.0  
5:00 212.4 F 0.0  
10:00 212.4 F 0.0  
15:00 197.4 F 0.0  
20:00 190.8 F 0.0  
25:00 187.3 F 0.0  
30:00 202.8 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 34:16  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:12:33  
CYCLE COMPLETE

08/31/2021 08:16 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_