
Midmark M11 - v1.0.5
Total Cycles: 589

Name: RITTER2
Sterilizer ID: V2313228
11/30/2021 01:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 109.4 F 0.0
2:00 108.6 F 0.0
4:00 118.8 F 0.0
6:00 148.7 F 0.2
8:00 189.3 F 2.8

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 196F
CHAMBER PRESS 5.0PSI
** ITEMS NOT STERILE **

11/30/2021 02:01 PM