

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4183

Name: RITTER 1  
Sterilizer ID: V2178198  
01/04/2024 03:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.8 F 0.0  
2:00 186.6 F 3.8  
4:00 215.5 F 9.5  
6:00 224.0 F 6.0  
8:00 235.9 F 10.8  
10:00 233.6 F 8.0  
12:00 241.6 F 10.7  
14:00 247.2 F 13.6  
16:00 259.7 F 20.4  
18:00 271.0 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.4 F 28.5  
2:00 273.0 F 28.9  
3:00 272.3 F 29.0  
4:00 272.2 F 28.4  
  
Min 271.6 F 28.3  
Max 273.0 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.4 F 0.0  
5:00 235.3 F 0.0  
10:00 234.2 F 0.0  
15:00 228.0 F 0.0  
20:00 227.3 F 0.0  
25:00 232.2 F 0.0  
30:00 220.8 F 0.0  
Drying Complete

FILLING: 2:19  
HEATING: 19:22  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 00:59:04  
CYCLE COMPLETE

01/04/2024 04:54 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_