
Midmark M11 - v1.0.0
Total Cycles: 3527

Name: RITTER 1
Sterilizer ID: V2178198
07/14/2023 09:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.0 F 0.0
2:00 189.0 F 0.6
4:00 197.9 F 4.2
6:00 211.9 F 8.8
8:00 225.5 F 12.6
10:00 230.1 F 8.3
12:00 241.5 F 13.3
14:00 240.6 F 11.0
16:00 237.7 F 9.3
18:00 251.6 F 15.9
20:00 262.8 F 22.3
22:00 272.8 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.6 F 28.9
2:00 271.9 F 28.3
3:00 271.6 F 28.4
4:00 272.7 F 28.9

Min 271.3 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 202.5 F 0.0
10:00 207.1 F 0.0
15:00 204.6 F 0.0
20:00 207.2 F 0.0
25:00 208.4 F 0.0
30:00 207.7 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 22:23
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:01:36
CYCLE COMPLETE

07/14/2023 10:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____