
Midmark M11 - v1.0.0
Total Cycles: 2670

Name: RITTER 1
Sterilizer ID: V2178198
12/13/2022 04:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.1 F 0.0
2:00 177.3 F 2.8
4:00 185.9 F 7.1
6:00 206.1 F 12.0
8:00 225.9 F 6.6
10:00 234.8 F 11.8
12:00 234.2 F 8.6
14:00 240.1 F 10.4
16:00 246.7 F 13.5
18:00 259.5 F 20.3
20:00 270.8 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.7 F 28.9
2:00 271.3 F 28.2
3:00 272.7 F 29.0
4:00 271.7 F 28.3

Min 271.3 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.4 F 0.0
5:00 220.4 F 0.0
10:00 226.3 F 0.0
15:00 226.9 F 0.0
20:00 236.2 F 0.0
25:00 234.8 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 20:47
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:43
CYCLE COMPLETE

12/13/2022 05:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____