
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 006012
11:49 19/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:03
STERILIZING 3:00
136.0 C 220kPa 3:00
136.1 C 222kPa 3:30
136.2 C 222kPa 4:00
136.0 C 222kPa 4:30
136.4 C 220kPa 5:00
136.4 C 222kPa 5:30
136.4 C 222kPa 6:00
136.4 C 224kPa 6:30
136.4 C 224kPa 6:30
VENTING 6:30
AIR DRYING 7:23
CYCLE COMPLETE 10:40

Digital Signature #
A5C8C00F562BC549
