
Midmark M11 - v1.0.0
Total Cycles: 1804

Name: RITTER 1
Sterilizer ID: V2178198
04/07/2022 10:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.9 F 0.0
2:00 122.9 F 0.1
4:00 164.2 F 2.7
6:00 176.1 F 6.4
8:00 207.5 F 10.7
10:00 218.1 F 3.8
12:00 228.6 F 8.6
14:00 239.1 F 13.4
16:00 238.0 F 10.2
18:00 236.3 F 8.3
20:00 248.7 F 14.6
22:00 260.6 F 21.2
24:00 271.3 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.4 F 28.8
2:00 271.8 F 28.8
3:00 272.1 F 28.5
4:00 272.3 F 29.1

Min 271.3 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.1 F 0.0
5:00 220.1 F 0.0
10:00 219.1 F 0.0
15:00 232.1 F 0.0
20:00 221.5 F 0.0
25:00 232.2 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 24:37
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:02:44
CYCLE COMPLETE

04/07/2022 11:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____