
Midmark M11 - v1.0.0
Total Cycles: 4660

Name: RITTER 1
Sterilizer ID: V2178198
06/19/2024 02:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.7 F 0.0
2:00 165.0 F 3.1
4:00 176.9 F 7.1
6:00 206.4 F 11.6
8:00 217.1 F 5.6
10:00 228.2 F 10.8
12:00 230.5 F 7.1
14:00 243.0 F 13.2
16:00 242.2 F 11.6
18:00 254.8 F 18.0
20:00 265.8 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 273.0 F 29.2
2:00 271.5 F 28.6
3:00 272.8 F 29.5
4:00 271.7 F 28.9

Min 271.0 F 28.4
Max 273.2 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.4 F 0.0
5:00 208.0 F 0.0
10:00 210.1 F 0.0
15:00 218.4 F 0.0
20:00 217.6 F 0.0
25:00 218.3 F 0.0
30:00 219.2 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 21:46
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:00:38
CYCLE COMPLETE

06/19/2024 03:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____