
Midmark M11 - v1.0.5
Total Cycles: 519

Name: RITTER2
Sterilizer ID: V2313228
11/08/2021 08:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.8 F 0.0
2:00 74.6 F 0.0
4:00 99.3 F 0.0
6:00 118.3 F 0.1
8:00 132.1 F 0.7
10:00 150.6 F 3.3
12:00 170.0 F 6.3
14:00 193.1 F 10.1
16:00 209.9 F 9.4
18:00 221.8 F 8.0
20:00 233.6 F 12.6
22:00 234.6 F 9.8
24:00 237.1 F 9.4
26:00 247.1 F 14.4
28:00 258.6 F 20.8
30:00 268.9 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.2 F 29.6
2:00 271.8 F 29.7
3:00 271.8 F 29.3
4:00 272.1 F 30.0

Min 271.3 F 29.1
Max 272.4 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 178.9 F 0.0
10:00 181.6 F 0.0
15:00 179.5 F 0.0
20:00 185.8 F 0.0
25:00 197.6 F 0.0
30:00 199.8 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 31:08
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:09:07
CYCLE COMPLETE

11/08/2021 09:52 AM
ACCEPTED: X
REJECTED:
Operator: U01