
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001650
15:03 27/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:40
STERILIZING 3:43
136.4 C 225kPa 3:43
136.2 C 224kPa 4:13
136.4 C 224kPa 4:43
136.4 C 225kPa 5:13
136.3 C 225kPa 5:43
136.2 C 223kPa 6:13
136.3 C 222kPa 6:43
136.2 C 223kPa 7:13
136.2 C 223kPa 7:13
VENTING 7:13
AIR DRYING 8:03
CYCLE COMPLETE 8:20

Digital Signature #
5216E1AD88BC35AD
