
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000673
11:16 21/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:32
PRESSURIZING 3:34
STERILIZING 4:41
136.4 C 223kPa 4:41
136.1 C 223kPa 5:11
136.3 C 224kPa 5:41
136.2 C 221kPa 6:11
136.2 C 221kPa 6:41
136.4 C 221kPa 7:11
136.2 C 222kPa 7:41
136.4 C 224kPa 8:11
136.4 C 224kPa 8:11
VENTING 8:11
AIR DRYING 9:06
CYCLE COMPLETE 9:10

Digital Signature #
C38237224F4F527F
