

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4152

Name: RITTER2  
Sterilizer ID: V2313228  
11/22/2024 09:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.1 F 0.0  
2:00 73.7 F 0.0  
4:00 106.5 F 0.0  
6:00 142.0 F 0.2  
8:00 166.0 F 0.4  
10:00 184.8 F 0.8  
12:00 194.3 F 2.5  
14:00 208.4 F 5.5  
16:00 220.3 F 9.0  
18:00 232.9 F 12.5  
20:00 228.7 F 7.6  
22:00 238.6 F 12.0  
24:00 235.6 F 9.3  
26:00 240.1 F 10.5  
28:00 245.3 F 13.3  
30:00 256.3 F 19.2  
32:00 266.4 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.5 F 29.5  
2:00 271.6 F 29.0  
3:00 272.5 F 29.5  
4:00 271.5 F 28.9  
  
Min 271.5 F 28.8  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.7 F 0.0  
5:00 200.1 F 0.0  
10:00 216.8 F 0.0  
15:00 212.7 F 0.0  
20:00 212.5 F 0.0  
25:00 212.8 F 0.0  
30:00 213.4 F 0.0  
  
Drying Complete

FILLING: 0:49  
HEATING: 33:44  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:11:56  
CYCLE COMPLETE

11/22/2024 11:04 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_