
Midmark M11 - v1.0.0
Total Cycles: 3520

Name: RITTER 1
Sterilizer ID: V2178198
07/12/2023 02:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.0 F 0.0
2:00 147.3 F 0.3
4:00 162.7 F 3.0
6:00 192.6 F 6.9
8:00 216.9 F 12.0
10:00 223.3 F 6.3
12:00 235.5 F 11.2
14:00 232.7 F 7.9
16:00 242.7 F 11.8
18:00 244.9 F 12.7
20:00 257.4 F 19.2
22:00 268.4 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.6 F 29.0
2:00 271.5 F 28.2
3:00 272.6 F 29.1
4:00 271.9 F 28.5

Min 271.3 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.5 F 0.0
5:00 207.8 F 0.0
10:00 215.8 F 0.0
15:00 226.3 F 0.0
20:00 226.6 F 0.0
25:00 230.2 F 0.0
30:00 231.0 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 23:16
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:25
CYCLE COMPLETE

07/12/2023 03:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____