
Midmark M11 - v1.0.5
Total Cycles: 4132

Name: RITTER2
Sterilizer ID: V2313228
11/18/2024 05:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.7 F 0.0
2:00 118.6 F 0.0
4:00 128.6 F 0.1
6:00 160.3 F 2.0
8:00 190.9 F 5.6
10:00 205.7 F 9.8
12:00 227.9 F 8.1
14:00 231.8 F 9.4
16:00 235.3 F 10.3
18:00 241.4 F 12.1
20:00 239.6 F 10.7
22:00 252.5 F 17.0
24:00 263.5 F 23.5
26:00 271.8 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.8 F 29.6
2:00 271.8 F 29.7
3:00 272.4 F 29.4
4:00 272.1 F 29.9

Min 271.2 F 29.1
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.1 F 0.0
5:00 244.3 F 0.0
10:00 229.0 F 0.0
15:00 219.1 F 0.0
20:00 230.8 F 0.0
25:00 226.9 F 0.0
30:00 206.3 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 26:53
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:05:15
CYCLE COMPLETE

11/18/2024 06:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____