
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001360
17:41 23/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:51
STERILIZING 3:55
136.4 C 220kPa 3:55
135.9 C 221kPa 4:25
136.0 C 222kPa 4:55
136.1 C 222kPa 5:25
136.4 C 225kPa 5:55
136.2 C 223kPa 6:25
136.3 C 224kPa 6:55
136.4 C 222kPa 7:25
136.4 C 222kPa 7:25
VENTING 7:25
AIR DRYING 8:18
CYCLE COMPLETE 14:45

Digital Signature #
DAB8A7CF4D845252
