
Midmark M11 - v1.0.0
Total Cycles: 4614

Name: RITTER 1
Sterilizer ID: V2178198
06/06/2024 01:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 96.6 F 0.0
2:00 97.6 F 0.7
4:00 120.7 F 2.8
6:00 141.0 F 5.3
8:00 166.8 F 8.0
10:00 190.6 F 11.3
12:00 207.7 F 3.5
14:00 212.4 F 7.1
16:00 225.6 F 10.9
18:00 227.3 F 7.0
20:00 237.8 F 11.5
22:00 235.1 F 8.9
24:00 247.8 F 14.7
26:00 258.4 F 20.5
28:00 268.3 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 271.5 F 29.3
2:00 273.0 F 30.0
3:00 271.9 F 29.3
4:00 272.9 F 30.0

Min 271.0 F 28.8
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.3 F 0.0
5:00 199.9 F 0.0
10:00 196.4 F 0.0
15:00 194.0 F 0.0
20:00 197.4 F 0.0
25:00 206.6 F 0.0
30:00 217.9 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 29:14
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:07:57
CYCLE COMPLETE

06/06/2024 02:25 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____